

UK Declaration of Conformity (DoC)

We manufacturer: Senbiosys SA
Route des Gouttes-d'Or 40
2000 Neuchâtel, Switzerland

declare that this DoC is issued under our sole responsibility for the equipment:

Velia Smart Ring

That the equipment above is in conformity with the following relevant legislation:

Radio Equipment Regulations 2017

The following harmonized standards and technical specifications have been applied:

Essential requirement	Standards
Regulation 6.1 (a) Safety	EN IEC 62368-1:2020+A11:2020
Regulation 6.1 (a) Health	EN 62479:2010 EN IEC 62311:2020
Regulation 6.1 (b) EMC	EN 55032:2015+A11:2020+A1:2020 EN 61000-3-3:2013+A1:2019+A2:2021 EN IEC 61000-3-2:2019+A1:2021+A2:2024 EN 55035:2017+A11:2020 EN 301 489-1 V2.2.3 EN 301 489-3 V2.3.2 EN 301 489-17 V3.3.1
Regulation 6.2 Radio Spectrum	EN 300 328 V2.2.2 EN 300 330 V2.1.1

Notified body: SGS North America, Inc. number NB 2906 performed Modules B - UKCA Type Examination Certificate and issued the UK Type Examination Certificates: AB2906.2025.000101 and AB2906.2025.000107.

RoHS statement:

The product complies with UK RoHS Regulations 2012 aligned with European Union's RoHS Directive 2011/65/EU. The product does not contain restricted substances in excess of the maximum concentration values tolerated by weight in homogeneous materials.

Standards and technical specifications applied: IEC 62321-3-1:2013, IEC 62321-7-1:2015, IEC 62321-6:2015, IEC 62321-8:2017

Signed for and on behalf of Senbiosys SA:

Neuchâtel, 26 September 2025


Taha BenMrad
Chief Executive Officer
Antonino Caizzone
Chief Engineering &
Operations Officer